

To protect the privacy of customers or suppliers of Salamander Jewelry. We erased their address from this document.

Document No. : TR-EO-005080217

Date : 8/2/2017

Page : 2 / 4

PRODUCT : ORDER NO. 2173732

### SAMPLE SLIP PARAMETER

EXHAUST:04H COMIN TOTAL :12H 12MIN			
PRETT :01H 06MIN STERI :07H COMIN			
ACTUAL TIME			
06:12:59 STERILIZATION SUCCESS			
END	06:12:48	056,50	-005kPa
	06:12:48	056,50	-005kPa
	02:16:50	056,70	-014kPa
	02:07:17	056,70	-001kPa
EXHAUST	02:05:47	056,80	000kPa
	01:57:17	056,70	000kPa
	01:47:17	056,80	000kPa
	01:37:17	056,80	000kPa
	01:27:17	056,70	-001kPa
	01:17:17	057,00	-001kPa
	01:07:17	057,00	000kPa
	00:57:17	057,10	000kPa
	00:47:17	057,00	001kPa
	00:37:17	057,00	000kPa
	00:27:17	056,90	000kPa
	00:17:17	056,90	-001kPa
	00:07:17	056,80	000kPa
	23:57:17	056,80	000kPa
	23:47:17	056,60	000kPa
	23:37:17	056,60	000kPa
	23:27:17	056,40	000kPa
	23:17:17	056,30	000kPa
	23:07:17	056,30	001kPa
	22:57:17	056,50	000kPa
	22:47:17	056,50	000kPa
	22:37:17	056,70	000kPa
	22:27:17	056,60	000kPa
	22:17:17	056,70	000kPa

	22:07:17	057,30	000kPa
	21:57:17	056,90	001kPa
	21:47:17	057,00	055kPa
	21:37:17	057,00	056kPa
	21:27:17	057,00	059kPa
	21:17:17	057,00	060kPa
	21:07:17	057,10	060kPa
	20:57:17	057,10	063kPa
	20:47:17	057,20	068kPa
	20:37:17	057,10	078kPa
	20:27:17	057,20	089kPa
	20:17:17	057,30	096kPa
	20:07:17	057,20	109kPa
	19:57:17	057,00	132kPa
	19:47:17	057,10	170kPa
	19:37:17	057,20	261kPa
	19:27:17	057,20	311kPa
	19:17:17	057,10	311kPa
	19:07:17	057,90	311kPa
STERI	19:06:32	057,30	311kPa
ADD EU	19:03:13	057,60	-088kPa
	18:57:17	057,40	-017kPa
HUMIDI	18:53:11	057,30	-086kPa
PRETT	18:47:17	057,10	000kPa
PREHEAT	18:00:55	050,90	000kPa
STATE	TIME	TEMP	PRE
RUN COUNT : 1079			
ED PRE	:060kPa	VENT TIME:04H COM	
STERI TEMP:	056,00	STERI TIME:07H COM	
PRO TIP	18:00:55	050,90	000kPa
START TIME: 2017-02-06 18:00:48			