

To protect the privacy of customers or suppliers of Salamander Jewelry. We erased their address from this document.

Document No. : TR-EO-001020317

Date : 2/3/2017

Page : 2 / 3

PRODUCT : ORDER NO. 2176513

SAMPLE SLIP PARAMETER

EXHAUST:04H (00MIN)	TOTAL :12H (00MIN)
PRETT :07H (00MIN)	STERI :07H (00MIN)
ACTUAL TIME	
05:39:23 STERILIZATION SUCCESS	
END	05:39:12 061,30 -001kPa
	05:39:12 061,30 -001kPa
	01:40:41 061,30 -015kPa
EXHAUST 40x06	
	01:30:41 061,40 000kPa
	01:20:41 061,30 000kPa
	01:10:41 061,30 001kPa
	01:00:41 061,30 000kPa
	00:50:41 061,40 001kPa
	00:40:41 061,30 001kPa
	00:30:41 061,30 000kPa
	00:20:41 061,30 -001kPa
	00:10:41 061,30 000kPa
	00:00:41 061,20 001kPa
	23:50:41 061,40 000kPa
	23:40:41 061,30 000kPa
	23:30:41 061,20 000kPa
	23:20:41 061,40 000kPa
	23:10:41 061,40 000kPa
	23:00:41 061,40 003kPa
	22:50:41 061,30 000kPa
	22:40:41 061,40 000kPa
	22:30:41 061,40 001kPa
	22:20:41 061,50 001kPa
	22:10:41 061,50 000kPa
	22:00:41 061,70 001kPa
	21:50:41 061,70 001kPa
	21:40:41 061,70 000kPa
	21:30:41 061,70 061kPa

	21:20:41 061,60 066kPa
	21:10:41 061,70 069kPa
	21:00:41 061,80 070kPa
	20:50:41 061,50 072kPa
	20:40:41 061,80 074kPa
	20:30:41 061,80 079kPa
	20:20:41 061,40 060kPa
	20:10:41 062,00 100kPa
	20:00:41 062,00 111kPa
	19:50:41 062,10 125kPa
	19:40:41 062,20 145kPa
	19:30:41 062,20 170kPa
	19:20:41 062,20 207kPa
	19:10:41 061,90 211kPa
	19:00:41 062,10 211kPa
	18:50:41 062,50 211kPa
	18:40:41 062,50 211kPa
STERI	18:34:10 062,00 211kPa
	18:20:41 062,00 -027kPa
	18:10:41 062,50 -030kPa
ADD EO	18:16:17 062,30 -060kPa
	18:10:41 062,30 -015kPa
HUMID1	18:06:34 062,10 -080kPa
PRETT	18:00:41 062,00 001kPa
PREHEAT	17:51:20 050,10 000kPa
STATE	TIME TEMP PRE
RUN COUNT : 1110	
EO PRE	:060kPa VENT TIME:04H 00M
STERI TEMP	:056,00 STERI TIME:07H 00M
PRE TIP	:17:31:20 050,10 000kPa
START TIME: 2017-02-28 17:31:20	