

Report Test

To protect the privacy of customers or suppliers of Salamander Jewelry.
We erased their address from this document.

LOT No : 0259/2021

Date : 7 July 2021

Determination of Sterilization Operation Specification

Item Detail	Specification	Setting	Result
Chamber Setting Temperature	40.0-55.5 °C	45.0 °C	
Vacuum pressure	40-60 kPa	60 kPa	
Vacuum Time	50-60 min	60 min	
Humidification			
Humidity	50.0-90.0%	20.2 %	
Sterilization			
Sterilization Temperature	40.0-55.5 °C	43.5 °C	
Gas Injection			
Final pressure	0-10 kPa	4 kPa	
Final temperature of vaporizer	40.0-55.5 °C	44.8 °C	
Exposure			
Temperature at start point of Exposure time period	40.0-55.5 °C	43.3 °C	
Temperature at finish point of Exposure time period	40.0-55.5 °C	46.0 °C	
Exposure Time	120-240 min	180 min	
Exhaust			
Exhaust Time	60-80 min	75 min	
First Vacuum pressure	0-2 kPa	1.5 kPa	
Second Vacuum pressure	0-2 kPa	1.5 kPa	
Third Vacuum pressure	0-2 kPa	1.5 kPa	
Fourth Vacuum pressure	0-2 kPa	1.5 kPa	
Fifth Vacuum pressure	0-2 kPa	1.5 kPa	
Storage			
Storage Time	≥ 48 hour	≥ 48 hour	
Chemical Test			
EO sterilization temperature 40-55°C, Time Sterilization by EO 3hours	Blue	Red	Pass
Biological Test			
EZ Test Biological Indicator	Green	Green	Pass

(.....*Jutha Liam*.....)
Mr. Jutha Liammukda
Research and Development Engineer
Salamander Jewelry Co., Ltd.

END

Air replacement:
Temperature: 44.9t
Humidity: 56.4RH
Pressures: -1.5KPa
Sterilization time: 0Hour
Air replace rounds: 1
round time: 0min(s)
2021-4-21 16:45

Air replacement:
Temperature: 44.9t
Humidity: 59.3RH
Pressures: -1.5KPa
Sterilization time: 0Hour
Air replace rounds: 2
round time: 0min(s)
2021-4-21 16:25

Air replacement:
Temperature: 44.9t
Humidity: 62.4RH
Pressures: -1.5KPa
Sterilization time: 0Hour
Air replace rounds: 3
round time: 0min(s)
2021-4-21 16:4

Air replacement:
Temperature: 44.9t
Humidity: 66.4RH
Pressures: -1.5KPa
Sterilization time: 0Hour
Air replace rounds: 4
round time: 0min(s)
2021-4-21 15:44

Air replacement:
Temperature: 46.0t
Humidity: 75.1RH
Pressures: -1.5KPa
Sterilization time: 0Hour
Air replace rounds: 5
round time: 15min(s)
2021-4-21 15:24

Sterilization:
Temperature: 45.6t
Humidity: 80.9RH
Pressures: -0.0KPa
Sterilization time: 0Hour
Air replace rounds: 5
round time: 15min(s)
2021-4-21 15:18

Sterilization:
Temperature: 43.3t
Humidity: 80.0RH
Pressures: -0.5KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2021-4-21 12:20

Sterilization:
Temperature: 43.7t
Humidity: 80.7RH
Pressures: -2.2KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2021-4-21 12:18

Add EO gas
Temperature: 43.6t
Humidity: 80.0RH
Pressures: -4.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2021-4-21 12:18

Evacuation
Temperature: 44.8t
Humidity: 71.4RH
Pressures: -0.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2021-4-21 12:12

Keeping Temp
Temperature: 45.1t
Humidity: 20.2RH
Pressures: -1.5KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2021-4-21 11:12

Keep pressure qualified
Temperature: 44.5t
Humidity: 20.2RH
Pressures: -59.2KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2021-4-21 11:11

Evacuation
Temperature: 42.7t
Humidity: 20.2RH
Pressures: -3.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2021-4-21 10:59

Heating-up:
Temperature: 23.9t
Humidity: 20.1RH
Pressures: -3.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2021-4-21 10:34

Setting:
Temperature: 45.0t
Humidity: 55.0RH
Pressures: -60.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2021-4-21 10:33



SALAMANDER
jewelry

Salamander Jewelry Co., Ltd.

128/1 Moo 7, Soi Wat Nakhon Chuenchum, Phutthamonthon Sai 4 Rd.
Tambon Krathumlom, Amphur Sampran, Nakhon Pathom 73220, Thailand.
Tel: 66-2814-4455 | Fax: 66-2-814-3300 | Email: info@body-piercing.com
Website: www.body-piercing.com

LINE	Code	Description	LOT	Color	BOX	TOTAL
2350587						
19	TDI-Z-TBNDJ-OP-1.6-14.0-5/8-AQ	STERILE TITANIUM JEWEL NAVEL BANANA FOR TDI	LOT-0259	AQ		25
20	TDI-Z-TBNDJ-OP-1.6-14.0-5/8-BZ	STERILE TITANIUM JEWEL NAVEL BANANA FOR TDI	LOT-0259	BZ		25
21	TDI-Z-TBNDJ-OP-1.6-14.0-5/8-RO	STERILE TITANIUM JEWEL NAVEL BANANA FOR TDI	LOT-0259	RO		25
22	TDI-Z-TBNJ-OP-1.6-10.0-5/8-AQ	STERILE TITANIUM JEWEL NAVEL BANANA FOR TDI	LOT-0259	AQ		100
49	TDI-Z-TMCB-1.2-10.0-3	STERILE TITANIUM MICRO CIRCULAR BARBELL FOR TDI	LOT-0259			300
52	TDI-Z-TMLB-1.2-10.0-3	STERILE TITANIUM MICRO LABRET FOR TDI	LOT-0259			300
54	TDI-Z-TMLB-1.2-8.0-2.5	STERILE TITANIUM MICRO LABRET FOR TDI	LOT-0259			300
55	TDI-Z-TMLB-1.2-8.0-3	STERILE TITANIUM MICRO LABRET FOR TDI	LOT-0259			700
56	TDI-Z-TMLB-1.2-8.0-2.5-CR	STERILE TITANIUM MICRO LABRET WITH JEWELLED BALL FOR TDI	LOT-0259	CR		300
61	TDI-Z-TNO-1.0-6.5-2.5	STERILE TITANIUM TITANIUM CURVED BALL NOSE STUD FOR TDI	LOT-0259			200
2356782						
1	PT-Z-TBCH-1.2-8.0-4	STERILE TITANIUM BALL CLOSURE RING WITH HEMATITE BALL FOR PTIT ROC	LOT-0259			115
5	PT-Z-TJNO-0.8-6.5-2.35-CR	STERILE TITANIUM JEWELLED CURVED NOSE STUD FOR PTIT ROC	LOT-0259	CR		110
8	PT-Z-TMBN-1.2-8.0-3	STERILE TITANIUM MICRO CURVED BARBELL FOR PTIT ROC	LOT-0259			128
9	PT-Z-TMCB-1.2-8.0-3	STERILE TITANIUM MICRO CIRCULAR BARBELL FOR PTIT ROC	LOT-0259			125
2351819						
7	THD-Z-CNOD-1.0-6.5-2.5	STERILE SURGICAL STEEL CURVED NOSE STUD FOR THENDRO	LOT-0259			160
26	THD-Z-TNO-1.0-6.5-2.5	STERILE SURGICAL STEEL 1.2MM THREADING CURVED NOSE STUD WITH BALL FOR THENDRO	LOT-0259			300
47	THD-Z-TMCB-1.2-10.0-3	STERILE TITANIUM MICRO CIRCULAR BARBELL FOR THENDRO	LOT-0259			700
2355414						
6	PT-Z-TMLB-1.2-8.0-3	STERILE TITANIUM MICRO LABRET FOR PTIT ROC	LOT-0259			390
7	PT-Z-TMLB-1.2-9.0-3	STERILE TITANIUM MICRO LABRET FOR PTIT ROC	LOT-0259			284
9	PT-Z-XMBN-1.1-2-8.0-3-CL	STERILE BIOFLEX MICRO CURVED BARBELL WITH STEEL BALLS FOR PTIT ROC	LOT-0259			52
2357623						
1	TDI-Z-TINMLB-1.2-10.0-3	STERILE TITANIUM INTERNALLY THREADED MICRO LABRET FOR TDI	LOT-0259			300
4	TDI-Z-TONR-1.0-7.0-2.5	STERILE TITANIUM OPEN NOSE RING FOR TDI	LOT-0259			100
7	TDI-Z-JNO90-0.8-6.5-2.35-CR	STERILE SURGICAL STEEL MICRO CURVED BARBELL FOR TDI	LOT-0259	CR		200
2352538						
22	PT-Z-TMCB-1.2-10.0-3	STERILE TITANIUM MICRO CIRCULAR BARBELL FOR PTIT ROC	LOT-0259			274
25	PT-Z-TMLB-1.2-8.0-3	STERILE TITANIUM MICRO LABRET FOR PTIT ROC	LOT-0259			916
2354764						
10	TDI-Z-TMLB-1.2-8.0-3	STERILE TITANIUM MICRO LABRET FOR TDI	LOT-0259			500
12	TDI-Z-TMLB-1.2-9.0-3	STERILE TITANIUM MICRO LABRET FOR TDI	LOT-0259			500
2351195						
1	QL-Z-BKFBC-1.0-9.0-3	STERILE BLACK PVD SURGICAL STEEL FIXED BEAD RING FOR QUILLIAN NO CARD	LOT-0259			50
4	QL-Z-FBC-1.0-7.0-3	STERILE SURGICAL STEEL FIXED BEAD RING FOR QUILLIAN NO CARD	LOT-0259			100
6	QL-Z-FBC-1.0-9.0-3	STERILE SURGICAL STEEL FIXED BEAD RING FOR QUILLIAN NO CARD	LOT-0259			100
2353247						
6	PT-Z-TBNDJ-1.6-12.0-5/8-CR	STERILE TITANIUM DOUBLE JEWEL NAVEL BANANA FOR PTIT ROC	LOT-0259	CR		148
12	PT-Z-TMLB-1.2-8.0-3	STERILE TITANIUM MICRO LABRET FOR PTIT ROC	LOT-0259			337
2351149						
13	PT-Z-TMCB-1.2-10.0-3	STERILE TITANIUM MICRO CIRCULAR BARBELL FOR PTIT ROC	LOT-0259			104
15	PT-Z-TMLB-1.2-10.0-3	STERILE TITANIUM MICRO LABRET FOR PTIT ROC	LOT-0259			264
2347052						
1	Z-PRTA-8.0-70.0	STERILE DISPOSABLE MULTI ANGLED RECEIVING TUBE 8.0MM	LOT-0259			543
2	Z-PRTA-6.0-70.0	STERILE DISPOSABLE MULTI ANGLED RECEIVING TUBE	LOT-0259			80
2354862						
13	PT-Z-TMCB-1.2-10.0-3	STERILE TITANIUM MICRO CIRCULAR BARBELL FOR PTIT ROC	LOT-0259			220
16	PT-Z-TMLB-1.2-8.0-3	STERILE TITANIUM MICRO LABRET FOR PTIT ROC	LOT-0259			582
2357886						
1	TDI-Z-MLB-1.2-8.0-3	STERILIZED SURGICAL STEEL MICRO LABRET FOR TDI	LOT-0259			300
2356635						
1	Z-TMBN-1.2-8.0-3-HP	STERILE TITANIUM CURVED MICRO BARBELL	LOT-0259			10
2349838						
7	TDI-Z-TMLB-1.2-10.0-2.5	STERILE TITANIUM MICRO LABRET FOR TDI	LOT-0259			300
2352456						
1	Z-GPJNO-1.0-6.5-2.35-CR	STERILE GOLD PLATED SURGICAL STEEL JEWELLED CURVED NOSE STUD	LOT-0259	CR		10
2357069						
1	Z-MLB-1.2-12.0-3	STERILE SURGICAL STEEL MICRO LABRET	LOT-0259			5

		2350044			
8	PT-Z-TMCB-1.2-10.0-3	STERILE TITANIUM MICRO CIRCULAR BARBELL FOR PTTT ROC	LOT-0259		144
		2353682			
1	TDI-Z-TMLB-1.2-8.0-3	STERILE TITANIUM MICRO LABRET FOR TDI	LOT-0259		300
		2355851			
1	Z-PNV-CRE	STERILE DISPOSAL ROUND CLAMPS	LOT-0259		200
		152076			
1	Z-GPINMLBJ-1.2-8.0-2-CR	STERILE GOLD PVD COATED SURGICAL STEEL INTERNALLY THREADED SWAROVSKI CRYSTAL JEWEL	LOT-0259	CR	10
2	Z-INMLBJ-1.2-6.0-1.8-CR	STERILE SURGICAL STEEL INTERNALLY THREADED SWAROVSKI CRYSTAL JEWEL MICRO LABRET	LOT-0259	CR	6
4	Z-SES1-3-CR	STERILE SURGICAL STEEL ROUND PRONG SET JEWEL EAR STUDS PAIR	LOT-0259	CR	5
5	Z-SES1-4-CHM	STERILE SURGICAL STEEL ROUND PRONG SET JEWEL EAR STUDS PAIR	LOT-0259	CHM	5
6	Z-SES1-4-CR	STERILE SURGICAL STEEL ROUND PRONG SET JEWEL EAR STUDS PAIR	LOT-0259	CR	5
7	Z-SES1-5-CR	STERILE SURGICAL STEEL ROUND PRONG SET JEWEL EAR STUDS PAIR	LOT-0259	CR	2
		152238			
1	Z-XUR-1.2-6X9-CL	STERILE BIOFLEX U SHAPE SEPTUM RETAINER	LOT-0259		1
2	Z-XINLBD2-1.2-6.0-2.5-CL	STERILE BIOFLEX INTERNAL LABRET WITH DISC	LOT-0259		1
		152166			
1	Z-BNJ-1.6-10.0-5/8-CR	STERILE SURGICAL STEEL JEWEL NAVEL BANANA	LOT-0259	CR	100
3	Z-MCB-1.2-10.0-3	STERILE SURGICAL STEEL MICRO CIRCULAR BARBELL	LOT-0259		100
		152154			
37	Z-SCRC008-1.2-8.0-CR	STERILE SURGICAL STEEL JEWEL ROOK CLICKER	LOT-0259	CR	3
38	Z-SCRC019-1.2-8.0-CR	STERILE SURGICAL STEEL PRONG SET JEWEL ROOK CLICKER	LOT-0259	CR	3
39	Z-SCRC022-1.2-8.0-CR	STERILE SURGICAL STEEL PRONG SET CRYSTAL JEWEL ROOK CLICKER	LOT-0259	CR	2
		150478			
23	Z-COBC-1.2-8.0	STERILE SURGICAL STEEL SEAMLESS RING	LOT-0259		20
		147049			
8	Z-TMCB-1.2-8.0-2.5-HP	STERILE TITANIUM MICRO CIRCULAR BARBELL	LOT-0259		10
		151944			
1	Z-TBL-1.6-16.0-5-HP	STERILE TITANIUM BARBELL	LOT-0259		50
2	Z-MBN-1.2-8.0-3	STERILE SURGICAL STEEL CURVED MICRO BARBELL	LOT-0259		200
		151943			
7	Z-MBN-1.2-12.0-3	STERILE SURGICAL STEEL CURVED MICRO BARBELL	LOT-0259		100
8	Z-MCB-1.2-8.0-3	STERILE SURGICAL STEEL MICRO CIRCULAR BARBELL	LOT-0259		200
		151947			
1	Z-BKBLDJS-1.6-14.0-5-CR	STERILE BLACK PVD COATED SURGICAL STEEL DOUBLE SIDE SWAROVSKI CRYSTALS JEWEL NIPP	LOT-0259	CR	2
		147049			
1	Z-TMCB-1.2-10.0-3-HP	STERILE TITANIUM MICRO CIRCULAR BARBELL	LOT-0259		10
		152187			
28	Z-XLB-3-1.6-10.0-3/5B-CL	STERILE BIOFLEX LABRET WITH TITANIUM BALL	LOT-0259		5
29	Z-XLB-3-1.6-12.0-3/5B-CL	STERILE BIOFLEX LABRET WITH TITANIUM BALL	LOT-0259		5
		152424			
1	Z-BCZ13-1.6-10.0-5-CR	STERILE RHODIUM PLATED BRASS JEWEL FLOWER NAVEL BANANA	LOT-0259	CR	2
		152298			
1	Z-SCSST7-1.2-6.0-CR	STERILE SURGICAL STEEL ROUND JEWEL HINGED SEPTUM CLICKER	LOT-0259	CR	4
2	Z-SCSST7-1.6-6.0-CR	STERILE SURGICAL STEEL ROUND JEWEL HINGED SEPTUM CLICKER	LOT-0259	CR	4
3	Z-MRAC-1.2-4X4	STERILE UV RASTA MICRO CONE	LOT-0259		4
4	Z-RGCOBC-0.8-10.0	STERILE ROSE GOLD PVD COATED SURGICAL STEEL SEAMLESS RING	LOT-0259		6
5	Z-RGCOBC-0.8-11.0	STERILE ROSE GOLD PVD COATED SURGICAL STEEL SEAMLESS RING	LOT-0259		4
6	Z-RGCOBC-0.8-6.0	STERILE ROSE GOLD PVD COATED SURGICAL STEEL SEAMLESS RING	LOT-0259		4
7	Z-RGCOBC-0.8-7.0	STERILE ROSE GOLD PVD COATED SURGICAL STEEL SEAMLESS RING	LOT-0259		4
8	Z-RGCOBC-0.8-8.0	STERILE ROSE GOLD PVD COATED SURGICAL STEEL SEAMLESS RING	LOT-0259		10
9	Z-RGCOBC-0.8-9.0	STERILE ROSE GOLD PVD COATED SURGICAL STEEL SEAMLESS RING	LOT-0259		6
10	Z-RGCOBC-1.0-10.0	STERILE ROSE GOLD PVD COATED SURGICAL STEEL SEAMLESS RING	LOT-0259		6
11	Z-RGCOBC-1.0-7.0	STERILE ROSE GOLD PVD COATED SURGICAL STEEL SEAMLESS RING	LOT-0259		6
		TOTAL			12131